Instrument Engineering Change Notice

Sakura Finetek Japan Co. Ltd.

NO. TEC6-2203

PRODUCT NAME	:	Tissue-Tek TEC6	DATE	:	June 15, 2022
PRODUCT CODES	:	5104, 5108, 5112, 5116, 5128	ORIGINATOR	:	Yoshizawa
IECN TYPE	:	PART DESIGN CHANGE	REFERENCE NUMBER	:	20027-1
SERIAL NUMBERS AFFECTED	:	See Remarks	ATTACHMENTS	:	N/A
FIELD RETROFIT REQUIRED	:	NO			

Title: Improved Position Adjustment of Dispenser Switch

[REASON AND DESCRIPTION FOR CHANGE]

The dispenser switch position adjustment has been improved by changing one of two slotted mounting screw holes to a circular hole.
With this change, the three related parts have been modified but they are not interchangeable with the old ones. Therefore, when any of the old parts needs to be replaced with the new one, the three parts need to be replaced as a set.

[PART CHANGE] **(IVDP=Inventory Disposal, ICHB=Inter Changeable)**

Description	Old	QTY	IVDP	New	QTY	ICHB
Micro Switch Mounting Plate	A-MAL2-0204-00-A	1		A-MAL2-0204-01-A	1	NO
Push Plate	A-MAL2-0205-00-A	1		A-MAL2-0205-01-A	1	NO
Dispenser Base	A-MAL2-0210-00-A	1		A-MAL2-0210-01-A	1	NO

[REMARKS]

Implementation:
J0: 5104 0324 and after
A1: 5108 0621 and after
E2: 5112 0429 and after
JC2: 5116 0165 and after
J2: 5128 0131 and after

[DISTRIBUTION]

- SFA ☑
- SFJ ☑
- SFE ☑
- SCN ☑

[APPROVAL]

Name : T. Sugawara

Signature :