Instrument Engineering Change Notice

NO. Prisma Plus-2202

Sakura Finetek Japan Co. Ltd.

PRODUCT NAME	: Tissue-Tek Prisma Plus	DATE	: June 15, 2022
PRODUCT CODES	: 6170, 6171, 6172, 6173, 6174, 6175, 6177	ORIGINATOR	: Y. Yazawa
		REFERENCE NUMBER	: 20014-2
IECN TYPE	: PART DESIGN CHANGE	ATTACHMENTS	: N/A
SERIAL NUMBERS AFFECTED	: See Remarks		
FIELD RETROFIT REQUIRED	: NO		

Title: Change of Circuit Protector

[REASON AND DESCRIPTION FOR CHANGE]

Due to product discontinuation, the circuit protector was replaced with the new part. Because the new circuit protector has different dimensions, a cutout was reshaped in the outer panel to avoid interfere with the new circuit protector.
Therefore, to keep interchangeability, the new circuit protector and outer panel need to be replaced together as a set.

[PART CHANGE] **(IVDP=Inventory Disposal, ICHB=Inter Changeable)**

Description	Old	QTY	IVDP	New	QTY	ICHB
Circuit Protector, 3120-F521-P7T1-W01D-15A	A3108042	1	NO	---	---	NO
Circuit Protector, 3120-N521-P7T1-W01D-15A	---	---		00001-4943	1	NO
Exterior (R-1)	A-MK25-0454-00-A	1	NO	A-MK25-0462-00-A	1	NO

[REMARKS]

AD/AS: from P.O. 206573
ED: 6172 0272 and after
ES: 6173 0365 and after
JD: 6174 0362 and after
JS: 6175 0079 and after
JCS: 6177 0270 and after

[DISTRIBUTION]

- SFA ☑
- SFJ ☑
- SFE ☑
- SCN ☑

[APPROVAL]

Name : T. Sugawara

Signature :